

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level1	

Lot No.	
53031401	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
RBC	10 ⁶ /μL	2.27	0.11	
HGB	g/dL	6.1	0.3	
	mmol/L	3.8	0.2	
HCT	%	17.5	1.8	
MCV	fL	77.1	4.6	
MCH	pg	26.9	2.4	
	fmol	1.668	0.150	
MCHC	g/dL	34.9	5.2	
	mmol/L	21.6	3.2	

Parameters		Target	Limit	Barcode
PLT-I	10 ³ /μL	54	22	
RDW-SD	fL	46.0	6.9	
RDW-CV	%	16.5	2.5	
PDW	fL	10.1	4.0	
PCT	%	0.06	0.03	
MPV	fL	10.9	1.6	
P-LCR	%	30.6	26.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level3	

Lot No.	
53031403	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
RBC	10 ⁶ /μL	5.28	0.26	
HGB	g/dL	16.8	0.7	
	mmol/L	10.4	0.4	
HCT	%	46.2	4.6	
MCV	fL	87.5	5.3	
MCH	pg	31.8	2.5	
	fmol	1.975	0.158	
MCHC	g/dL	36.4	5.1	
	mmol/L	22.6	3.2	

Parameters		Target	Limit	Barcode
PLT-I	10 ³ /μL	555	72	
RDW-SD	fL	44.6	6.7	
RDW-CV	%	13.9	2.1	
PDW	fL	10.4	2.1	
PCT	%	0.57	0.14	
MPV	fL	10.3	1.5	
P-LCR	%	25.5	12.8	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level3	

Lot No.	
53031403	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
WBC-C	10 ³ /μL	16.31	1.47	
WBC-D	10 ³ /μL	16.28	1.30	
NEUT#	10 ³ /μL	8.10	2.43	
LYMPH#	10 ³ /μL	3.32	1.00	
MONO#	10 ³ /μL	1.71	1.54	
EO#	10 ³ /μL	2.11	1.06	
BASO#	10 ³ /μL	1.04	0.81	


Parameters		Target	Limit	Barcode
IG#	10 ³ /μL	2.23	0.56	
NEUT%	%	49.8	7.5	
LYMPH%	%	20.4	4.1	
MONO%	%	10.5	5.3	
EO%	%	12.9	6.5	
BASO%	%	6.4	5.0	
IG%	%	13.7	3.4	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level3	

Lot No.	
53031403	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
RET#	10 ⁶ /μL	0.0464	0.0139	
RET%	%	0.88	0.31	
RBC-O	10 ⁶ /μL	5.11	0.51	
HFR	%	1.9	11.0	
MFR	%	13.6	30.0	
LFR	%	84.5	38.0	

Parameters		Target	Limit	Barcode
IRF	%	15.5	38.0	
RET-He	pg	26.0	2.6	
	fmol	1.614	0.161	
PLT-O	10 ³ /μL	566	85	
IPF#	10 ³ /μL	101.7	71.2	
IPF	%	18.0	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level2	

Lot No.	
53031402	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
RBC	10 ⁶ /μL	4.21	0.21	
HGB	g/dL	12.5	0.5	
	mmol/L	7.8	0.3	
HCT	%	35.2	3.5	
MCV	fL	83.5	5.0	
MCH	pg	29.7	2.4	
	fmol	1.843	0.147	
MCHC	g/dL	35.5	5.0	
	mmol/L	22.0	3.1	


Parameters		Target	Limit	Barcode
PLT-I	10 ³ /μL	237	36	
RDW-SD	fL	43.4	6.5	
RDW-CV	%	14.2	2.1	
PDW	fL	10.2	3.1	
PCT	%	0.25	0.08	
MPV	fL	10.4	1.6	
P-LCR	%	26.5	13.3	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level2	

Lot No.	
53031402	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
WBC-C	10 ³ /μL	7.20	0.65	
WBC-D	10 ³ /μL	7.19	0.58	
NEUT#	10 ³ /μL	3.37	1.01	
LYMPH#	10 ³ /μL	1.87	0.56	
MONO#	10 ³ /μL	0.60	0.54	
EO#	10 ³ /μL	0.90	0.45	
BASO#	10 ³ /μL	0.45	0.35	

Parameters		Target	Limit	Barcode
IG#	10 ³ /μL	0.92	0.28	
NEUT%	%	47.0	7.1	
LYMPH%	%	26.1	5.2	
MONO%	%	8.3	5.0	
EO%	%	12.4	6.2	
BASO%	%	6.2	4.8	
IG%	%	12.9	3.9	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level2	

Lot No.	
53031402	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
RET#	10 ⁶ /μL	0.0966	0.0290	
RET%	%	2.29	0.46	
RBC-O	10 ⁶ /μL	4.09	0.41	
HFR	%	2.8	11.0	
MFR	%	17.4	30.0	
LFR	%	79.8	38.0	


Parameters		Target	Limit	Barcode
IRF	%	20.2	38.0	
RET-He	pg	24.0	2.4	
	fmol	1.489	0.149	
PLT-O	10 ³ /μL	245	74	
IPF#	10 ³ /μL	44.6	33.5	
IPF	%	18.2	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level1	

Lot No.	
53031401	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
WBC-C	10 ³ /μL	2.51	0.25	
WBC-D	10 ³ /μL	2.52	0.25	
NEUT#	10 ³ /μL	1.15	0.35	
LYMPH#	10 ³ /μL	0.75	0.30	
MONO#	10 ³ /μL	0.18	0.16	
EO#	10 ³ /μL	0.29	0.15	
BASO#	10 ³ /μL	0.15	0.12	

Parameters		Target	Limit	Barcode
IG#	10 ³ /μL	0.31	0.09	
NEUT%	%	45.4	9.1	
LYMPH%	%	29.8	11.9	
MONO%	%	7.2	5.8	
EO%	%	11.7	5.9	
BASO%	%	5.9	4.6	
IG%	%	12.2	3.7	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level1	

Lot No.	
53031401	

Exp. Date	
2016/02/09	

Parameters		Target	Limit	Barcode
RET#	10 ⁶ /μL	0.1198	0.0359	
RET%	%	5.28	1.06	
RBC-O	10 ⁶ /μL	2.23	0.22	
HFR	%	2.2	11.0	
MFR	%	14.9	30.0	
LFR	%	82.9	38.0	

Parameters		Target	Limit	Barcode
IRF	%	17.1	38.0	
RET-He	pg	23.1	2.3	
	fmol	1.434	0.143	
PLT-O	10 ³ /μL	57	40	
IPF#	10 ³ /μL	10.3	9.8	
IPF	%	18.2	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.