< *e*-CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 61860804 Exp. date 2016-09-25 **L1 : Level 1**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	2.31	5.8	3.6	17.7	76.8	25.1	1.56	32.8	20.3	60
LIMIT	0.12	0.2	0.1	1.1	3.8	1.3	0.08	2.0	1.2	24

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	46.5	16.8	8.3	0.06	9.5	14.7
LIMIT	7.0	2.5	3.3	0.03	1.4	12.5

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	3.02	2.98	1.09	1.17	0.34	0.26	0.16
LIMIT	0.30	0.30	0.33	0.47	0.31	0.13	0.12

Please refer to the help message provided in the assay CD-ROM for the handling of the *e*-CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915

< *e*-CHECK(XS) ASSAY SHEET >

For XS-Series

Lot No. 61860805 Exp. date 2016-09-25 **L2 : Level 2**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	4.37	11.9	7.4	35.0	80.2	27.2	1.69	34.0	21.1	228
LIMIT	0.22	0.4	0.2	1.8	4.0	1.4	0.08	2.0	1.3	34

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	44.6	15.7	8.9	0.23	10.0	16.7
LIMIT	6.7	2.4	2.7	0.07	1.5	8.4

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	6.98	6.89	2.66	2.47	0.77	0.67	0.41
LIMIT	0.42	0.41	0.80	0.74	0.69	0.34	0.32

Please refer to the help message provided in the assay CD-ROM for the handling of the *e*-CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915