



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level1	


Lot No.	Barcode
61901401	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
RBC	$10^6/\mu\text{L}$	2.22	0.11	
HGB	g/dL	5.6	0.3	
	mmol/L	3.5	0.2	
HCT	%	16.2	1.6	
MCV	fL	72.8	4.4	
MCH	pg	25.2	2.3	
	fmol	1.565	0.141	
MCHC	g/dL	34.6	5.2	
	mmol/L	21.5	3.2	

Parameters		Target	Limit	Barcode
PLT-I	$10^3/\mu\text{L}$	57	23	
RDW-SD	fL	46.0	6.9	
RDW-CV	%	17.3	2.6	
PDW	fL	8.8	3.5	
PCT	%	0.06	0.03	
MPV	fL	10.4	1.6	
P-LCR	%	24.1	20.5	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level1	

Lot No.	Barcode
61901401	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
WBC-C	10 ³ /µL	2.50	0.25	
WBC-D	10 ³ /µL	2.49	0.25	
NEUT#	10 ³ /µL	1.14	0.34	
LYMPH#	10 ³ /µL	0.70	0.28	
MONO#	10 ³ /µL	0.21	0.19	
EO#	10 ³ /µL	0.29	0.15	
BASO#	10 ³ /µL	0.15	0.12	
IG#	10 ³ /µL	0.32	0.10	
AS-LYMP#	10 ³ /µL	0.12	0.07	

Parameters		Target	Limit	Barcode
NEUT%	%	45.5	9.1	
LYMPH%	%	28.2	11.3	
MONO%	%	8.4	6.7	
EO%	%	11.8	5.9	
BASO%	%	6.1	4.8	
IG%	%	12.8	3.8	
AS-LYMP%	%	4.9	2.9	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level1	

Lot No.	Barcode
61901401	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
RET#	$10^6/\mu\text{L}$	0.1191	0.0357	
RET%	%	5.36	1.07	
RBC-O	$10^6/\mu\text{L}$	2.17	0.22	
HFR	%	1.0	11.0	
MFR	%	12.0	30.0	
LFR	%	87.0	38.0	
IRF	%	13.0	38.0	
NEUT-RI	FI	60.4	24.2	
NEUT-GI	SI	209.0	20.9	


Parameters		Target	Limit	Barcode
RET-He	pg	22.1	2.2	
	fmol	1.372	0.137	
RBC-He	pg	24.5	4.9	
	fmol	1.520	0.304	
PLT-O	$10^3/\mu\text{L}$	59	41	
IPF#	$10^3/\mu\text{L}$	10.5	10.0	
IPF	%	17.9	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET


For XN-L series

Material		
XN-L Control Level2		

Lot No.		
61901402		

Exp. Date			
2016/10/18			

Parameters		Target	Limit	Barcode
RBC	$10^6/\mu\text{L}$	4.43	0.22	
HGB	g/dL	12.5	0.5	
	mmol/L	7.8	0.3	
HCT	%	35.2	3.5	
MCV	fL	79.4	4.8	
MCH	pg	28.2	2.3	
	fmol	1.751	0.140	
MCHC	g/dL	35.5	5.0	
	mmol/L	22.0	3.1	

Parameters		Target	Limit	Barcode
PLT-I	$10^3/\mu\text{L}$	222	33	
RDW-SD	fL	48.3	7.2	
RDW-CV	%	16.7	2.5	
PDW	fL	9.0	2.7	
PCT	%	0.22	0.07	
MPV	fL	10.2	1.5	
P-LCR	%	22.1	11.1	

There are some parameters which is not displayed on IPU according to the instrument.
 Please refer to the package insert for the handling of the XN-L CHECK.
 Do not leave control blood in the room temperature over an hour.
 Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level2	

Lot No.	Barcode
61901402	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
WBC-C	10 ³ /µL	6.80	0.61	
WBC-D	10 ³ /µL	6.79	0.54	
NEUT#	10 ³ /µL	3.25	0.98	
LYMPH#	10 ³ /µL	1.69	0.51	
MONO#	10 ³ /µL	0.58	0.52	
EO#	10 ³ /µL	0.84	0.42	
BASO#	10 ³ /µL	0.43	0.34	
IG#	10 ³ /µL	0.88	0.26	
AS-LYMP#	10 ³ /µL	0.27	0.14	

Parameters		Target	Limit	Barcode
NEUT%	%	47.9	7.2	
LYMPH%	%	24.9	5.0	
MONO%	%	8.5	5.1	
EO%	%	12.3	6.2	
BASO%	%	6.4	5.0	
IG%	%	13.0	3.9	
AS-LYMP%	%	4.0	2.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level2	


Lot No.	Barcode
61901402	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
RET#	$10^6/\mu\text{L}$	0.1009	0.0303	
RET%	%	2.28	0.46	
RBC-O	$10^6/\mu\text{L}$	4.30	0.43	
HFR	%	0.3	11.0	
MFR	%	7.5	30.0	
LFR	%	92.2	38.0	
IRF	%	7.8	38.0	
NEUT-RI	FI	59.2	23.7	
NEUT-GI	SI	209.9	21.0	

Parameters		Target	Limit	Barcode
RET-He	pg	22.7	2.3	
	fmol	1.409	0.141	
RBC-He	pg	27.4	5.5	
	fmol	1.700	0.340	
PLT-O	$10^3/\mu\text{L}$	228	68	
IPF#	$10^3/\mu\text{L}$	40.8	30.6	
IPF	%	17.9	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level3	

Lot No.	Barcode
61901403	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
RBC	$10^6/\mu\text{L}$	5.27	0.26	
HGB	g/dL	16.5	0.7	
	mmol/L	10.2	0.4	
HCT	%	45.5	4.6	
MCV	fL	86.4	5.2	
MCH	pg	31.3	2.5	
	fmol	1.943	0.155	
MCHC	g/dL	36.3	5.1	
	mmol/L	22.5	3.2	

Parameters		Target	Limit	Barcode
PLT-I	$10^3/\mu\text{L}$	553	72	
RDW-SD	fL	44.2	6.6	
RDW-CV	%	14.1	2.1	
PDW	fL	9.3	1.9	
PCT	%	0.55	0.14	
MPV	fL	9.9	1.5	
P-LCR	%	20.6	10.3	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	Barcode
XN-L Control Level3	


Lot No.	Barcode
61901403	

Exp. Date	Barcode
2016/10/18	

Parameters		Target	Limit	Barcode
WBC-C	10 ³ /µL	16.23	1.46	
WBC-D	10 ³ /µL	16.20	1.30	
NEUT#	10 ³ /µL	8.48	2.54	
LYMPH#	10 ³ /µL	3.05	0.92	
MONO#	10 ³ /µL	1.37	1.23	
EO#	10 ³ /µL	2.15	1.08	
BASO#	10 ³ /µL	1.15	0.90	
IG#	10 ³ /µL	2.40	0.60	
AS-LYMP#	10 ³ /µL	0.46	0.23	

Parameters		Target	Limit	Barcode
NEUT%	%	52.3	7.8	
LYMPH%	%	18.8	3.8	
MONO%	%	8.5	4.3	
EO%	%	13.3	6.7	
BASO%	%	7.1	5.5	
IG%	%	14.8	3.7	
AS-LYMP%	%	2.8	1.4	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material			
XN-L Control Level3			

Lot No.			
61901403			

Exp. Date			
2016/10/18			

Parameters		Target	Limit	Barcode
RET#	$10^6/\mu\text{L}$	0.0471	0.0141	
RET%	%	0.89	0.31	
RBC-O	$10^6/\mu\text{L}$	5.10	0.51	
HFR	%	0.1	11.0	
MFR	%	6.5	30.0	
LFR	%	93.4	38.0	
IRF	%	6.6	38.0	
NEUT-RI	FI	54.4	21.8	
NEUT-GI	SI	209.1	20.9	

Parameters		Target	Limit	Barcode
RET-He	pg	24.5	2.5	
	fmol	1.520	0.152	
RBC-He	pg	30.5	6.1	
	fmol	1.893	0.379	
PLT-O	$10^3/\mu\text{L}$	555	83	
IPF#	$10^3/\mu\text{L}$	97.0	67.9	
IPF	%	17.5	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.