







XN-L CHECK™ ASSAY SHEET



For XN-L series

Material	
XN-L Control Level1	

Exp. Date	
2017/04/04	

Parameters		Target	Limit	Barcode
RBC	10 ⁶ /μL	2.33	0.12	
HGB	g/dL	6.1	0.3	
	mmol/L	3.8	0.2	
HCT	%	17.6	1.8	
MCV	fL	75.6	4.5	
MCH	pg	26.2	2.4	
	fmol	1.625	0.146	
MCHC	g/dL	34.7	5.2	
	mmol/L	21.5	3.2	


Lot No.	
63581401	


Parameters		Target	Limit	Barcode
PLT-I	10 ³ /μL	54	22	
RDW-SD	fL	47.4	7.1	
RDW-CV	%	17.1	2.6	
PDW	fL	9.2	3.7	
PCT	%	0.06	0.03	
MPV	fL	10.3	1.5	
P-LCR	%	23.5	20.0	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET




For XN-L series

Material	
XN-L Control Level1	

Lot No.	
63581401	

Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
WBC-C	10 ³ /μL	2.36	0.24	
WBC-D	10 ³ /μL	2.35	0.24	
NEUT#	10 ³ /μL	1.09	0.33	
LYMPH#	10 ³ /μL	0.63	0.25	
MONO#	10 ³ /μL	0.20	0.18	
EO#	10 ³ /μL	0.28	0.14	
BASO#	10 ³ /μL	0.15	0.12	
IG#	10 ³ /μL	0.30	0.09	
AS-LYMP#	10 ³ /μL	0.11	0.07	


Parameters		Target	Limit	Barcode
NEUT%	%	46.2	9.2	
LYMPH%	%	27.0	10.8	
MONO%	%	8.4	6.7	
EO%	%	12.0	6.0	
BASO%	%	6.4	5.0	
IG%	%	12.7	3.8	
AS-LYMP%	%	4.5	2.7	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level1	

Lot No.	
63581401	

Exp. Date	
2017/04/04	

Parameters		Target	Limit	Barcode
RET#	10 ⁶ /μL	0.1259	0.0378	
RET%	%	5.40	1.08	
RBC-O	10 ⁶ /μL	2.27	0.23	
HFR	%	1.5	11.0	
MFR	%	13.7	30.0	
LFR	%	84.8	38.0	
IRF	%	15.2	38.0	
NEUT-RI	FI	64.4	25.8	
NEUT-GI	SI	204.7	20.5	

Parameters		Target	Limit	Barcode
RET-He	pg	22.7	2.3	
	fmol	1.409	0.141	
RBC-He	pg	25.2	5.0	
	fmol	1.564	0.313	
PLT-O	10 ³ /μL	62	43	
IPF#	10 ³ /μL	11.0	10.5	
IPF	%	17.7	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level1	

Lot No.	
63581401	


Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
HYPO-He	%	1.5	2.0	
HYPER-He	%	0.3	0.3	
MicroR	%	14.4	11.5	
MacroR	%	2.2	1.3	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.





XN-L CHECK™ ASSAY SHEET



For XN-L series

Material	
XN-L Control Level2	

Lot No.	
63581402	

Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
RBC	10 ⁶ /μL	4.42	0.22	
HGB	g/dL	12.5	0.5	
	mmol/L	7.8	0.3	
HCT	%	35.1	3.5	
MCV	fL	79.5	4.8	
MCH	pg	28.3	2.3	
	fmol	1.755	0.140	
MCHC	g/dL	35.6	5.0	
	mmol/L	22.1	3.1	


Parameters		Target	Limit	Barcode
PLT-I	10 ³ /μL	232	35	
RDW-SD	fL	44.0	6.6	
RDW-CV	%	15.2	2.3	
PDW	fL	9.2	2.8	
PCT	%	0.23	0.07	
MPV	fL	9.8	1.5	
P-LCR	%	19.8	9.9	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET


For XN-L series

Material	
XN-L Control Level2	

Lot No.	
63581402	

Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
WBC-C	10 ³ /μL	6.78	0.61	
WBC-D	10 ³ /μL	6.77	0.54	
NEUT#	10 ³ /μL	3.40	1.02	
LYMPH#	10 ³ /μL	1.49	0.45	
MONO#	10 ³ /μL	0.54	0.49	
EO#	10 ³ /μL	0.88	0.44	
BASO#	10 ³ /μL	0.46	0.36	
IG#	10 ³ /μL	0.95	0.29	
AS-LYMP#	10 ³ /μL	0.24	0.12	


Parameters		Target	Limit	Barcode
NEUT%	%	50.3	7.5	
LYMPH%	%	22.0	4.4	
MONO%	%	7.9	4.7	
EO%	%	13.0	6.5	
BASO%	%	6.8	5.3	
IG%	%	14.1	4.2	
AS-LYMP%	%	3.6	1.8	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.

XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level2	

Lot No.	
63581402	

Exp. Date	
2017/04/04	

Parameters		Target	Limit	Barcode
RET#	10 ⁶ /μL	0.1146	0.0344	
RET%	%	2.59	0.52	
RBC-O	10 ⁶ /μL	4.27	0.43	
HFR	%	2.1	11.0	
MFR	%	16.6	30.0	
LFR	%	81.3	38.0	
IRF	%	18.7	38.0	
NEUT-RI	FI	62.1	24.8	
NEUT-GI	SI	204.3	20.4	

Parameters		Target	Limit	Barcode
RET-He	pg	23.3	2.3	
	fmol	1.446	0.145	
RBC-He	pg	27.3	5.5	
	fmol	1.694	0.339	
PLT-O	10 ³ /μL	250	75	
IPF#	10 ³ /μL	44.5	33.4	
IPF	%	17.8	5.0	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level2	

Lot No.	
63581402	


Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
HYPO-He	%	0.3	1.0	
HYPER-He	%	0.7	0.7	
MicroR	%	6.7	5.4	
MacroR	%	4.1	2.1	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.








XN-L CHECK™ ASSAY SHEET






For XN-L series

Material	
XN-L Control Level3	

Lot No.	
63581403	

Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
RBC	10 ⁶ /μL	5.38	0.27	
HGB	g/dL	16.9	0.7	
	mmol/L	10.5	0.4	
HCT	%	46.0	4.6	
MCV	fL	85.5	5.1	
MCH	pg	31.4	2.5	
	fmoI	1.949	0.156	
MCHC	g/dL	36.7	5.1	
	mmol/L	22.8	3.2	


Parameters		Target	Limit	Barcode
PLT-I	10 ³ /μL	566	74	
RDW-SD	fL	44.0	6.6	
RDW-CV	%	14.3	2.1	
PDW	fL	9.5	1.9	
PCT	%	0.54	0.14	
MPV	fL	9.6	1.4	
P-LCR	%	18.7	9.4	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.






XN-L CHECK™ ASSAY SHEET






For XN-L series

Material	
XN-L Control Level3	

Lot No.	
63581403	

Exp. Date	
2017/04/04	


Parameters		Target	Limit	Barcode
WBC-C	10 ³ /μL	15.76	1.42	
WBC-D	10 ³ /μL	15.73	1.26	
NEUT#	10 ³ /μL	8.22	2.47	
LYMPH#	10 ³ /μL	2.94	0.88	
MONO#	10 ³ /μL	1.33	1.20	
EO#	10 ³ /μL	2.15	1.08	
BASO#	10 ³ /μL	1.09	0.85	
IG#	10 ³ /μL	2.30	0.58	
AS-LYMP#	10 ³ /μL	0.45	0.23	


Parameters		Target	Limit	Barcode
NEUT%	%	52.4	7.9	
LYMPH%	%	18.7	3.7	
MONO%	%	8.4	4.2	
EO%	%	13.6	6.8	
BASO%	%	6.9	5.4	
IG%	%	14.6	3.7	
AS-LYMP%	%	2.9	1.5	





There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.


XN-L CHECK™ ASSAY SHEET



For XN-L series

Material	
XN-L Control Level3	

Exp. Date	
2017/04/04	

Parameters		Target	Limit	Barcode
RET#	10 ⁶ /μL	0.0505	0.0152	
RET%	%	0.94	0.33	
RBC-O	10 ⁶ /μL	5.23	0.52	
HFR	%	1.7	11.0	
MFR	%	13.2	30.0	
LFR	%	85.1	38.0	
IRF	%	14.9	38.0	
NEUT-RI	FI	59.1	23.6	
NEUT-GI	SI	203.9	20.4	


Lot No.	
63581403	


Parameters		Target	Limit	Barcode
RET-He	pg	24.1	2.4	
	fmol	1.496	0.150	
RBC-He	pg	30.4	6.1	
	fmol	1.887	0.377	
PLT-O	10 ³ /μL	593	89	
IPF#	10 ³ /μL	104.2	72.9	
IPF	%	17.6	5.0	


There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.



XN-L CHECK™ ASSAY SHEET

For XN-L series

Material	
XN-L Control Level3	

Lot No.	
63581403	

Exp. Date	
2017/04/04	

Parameters		Target	Limit	Barcode
HYPO-He	%	0.1	0.5	
HYPER-He	%	1.1	1.0	
MicroR	%	2.2	2.2	
MacroR	%	4.7	2.4	

There are some parameters which is not displayed on IPU according to the instrument.
Please refer to the package insert for the handling of the XN-L CHECK.
Do not leave control blood in the room temperature over an hour.
Please store it in a refrigerator (2-8°C) immediately after use.