

< e-CHECK[®](XE) ASSAY SHEET >

For XT-Series

Lot No. 70220810

Exp. date 2017-04-16

L1 : Level 1

MODE		RBC 10 ⁶ /uL	HGB		HCT %	MCV fL	MCH		MCHC		PLT 10 ³ /uL
			g/dL	mmol/L			pg	fmol	g/dL	mmol/L	
Open	TARGET	2.31	5.9	3.7	17.7	76.7	25.5	1.60	33.3	20.9	58
	LIMIT	0.12	0.2	0.1	1.1	3.8	1.3	0.08	2.0	1.3	23
Closed	TARGET	2.33	6.0	3.7	17.8	76.5	25.8	1.59	33.7	20.8	56
	LIMIT	0.12	0.2	0.1	1.1	3.8	1.3	0.08	2.0	1.2	22

MODE		RDW-SD fL	RDW-CV %	PDW fL	PCT %	MPV fL	P-LCR %
Open	TARGET	44.8	16.3	7.6	0.05	9.2	13.9
	LIMIT	6.7	2.4	1.5	0.03	1.4	11.8
Closed	TARGET	44.7	16.3	7.8	0.05	9.1	13.8
	LIMIT	6.7	2.4	1.6	0.03	1.4	11.7

MODE		WBC 10 ³ /uL	WBC-D ^{*1} 10 ³ /uL	NEUT# 10 ³ /uL	LYMPH# 10 ³ /uL	MONO# 10 ³ /uL	EO# 10 ³ /uL	BASO# 10 ³ /uL	PLT-O 10 ³ /uL
Open	TARGET	3.01	2.74	1.32	1.03	0.38	0.28	1.81	54
	LIMIT	0.30	0.27	0.40	0.41	0.34	0.14	1.50	24
Closed	TARGET	2.94	2.78	1.31	1.00	0.36	0.27	1.78	55
	LIMIT	0.29	0.28	0.39	0.40	0.32	0.14	1.45	25

MODE		RET# 10 ⁶ /uL	RET% %	RBC-O 10 ⁶ /uL	HFR %	MFR %	LFR %	IRF %
Open	TARGET	0.1208	5.23	2.43	3.1	21.0	75.9	24.1
	LIMIT	0.0362	1.57	0.24	8.0	28.0	33.0	36.0
Closed	TARGET	0.1107	4.75	2.51	3.0	20.8	76.2	23.8
	LIMIT	0.0332	1.43	0.25	8.0	28.0	33.0	36.0

MODE		IG# ^{*2} 10 ³ /uL	RET-He ^{*3}	
			pg	fmol
Open	TARGET	0.34	23.7	1.47
	LIMIT	0.10	2.4	0.15
Closed	TARGET	0.35	23.7	1.47
	LIMIT	0.11	2.4	0.15

*1 : Available only on XT-4000i.

*2 : Available only on XT-4000i, XT-2000i, or XT-1800i installed the optional software, XT IG master.

*3 : Available only on XT-4000i, or XT-2000i installed the optional software, XT RET master.

Please refer to the help message provided in the assay CD for the handling of the e-CHECK.(XE).

Please make sure the lot number before reading the data from the assay CD.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915

< e-CHECK[®](XE) ASSAY SHEET >

For XT-Series

Lot No. 70230811

Exp. date 2017-04-16

L2 : Level 2

MODE		RBC	HGB		HCT	MCV	MCH		MCHC		PLT
		10 ⁶ /uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10 ³ /uL
Open	TARGET	4.41	12.3	7.6	35.9	81.5	27.9	1.72	34.3	21.2	229
	LIMIT	0.22	0.4	0.2	1.8	4.1	1.4	0.09	2.1	1.3	34
Closed	TARGET	4.44	12.5	7.8	36.1	81.2	28.2	1.76	34.6	21.6	224
	LIMIT	0.22	0.4	0.2	1.8	4.1	1.4	0.09	2.1	1.3	34

MODE		RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
		fL	%	fL	%	fL	%
Open	TARGET	44.7	15.4	8.2	0.22	9.5	15.4
	LIMIT	6.7	2.3	1.6	0.07	1.4	7.7
Closed	TARGET	44.3	15.3	8.3	0.21	9.5	15.3
	LIMIT	6.6	2.3	1.7	0.06	1.4	7.7

MODE		WBC	WBC-D ^{†1}	NEUT#	LYMPH#	MONO#	EO#	BASO#	PLT-O
		10 ³ /uL	10 ³ /uL	10 ³ /uL	10 ³ /uL	10 ³ /uL	10 ³ /uL	10 ³ /uL	10 ³ /uL
Open	TARGET	7.29	6.38	3.57	2.17	0.81	0.74	4.69	217
	LIMIT	0.44	0.38	1.07	0.65	0.73	0.37	3.66	33
Closed	TARGET	7.11	6.65	3.49	2.19	0.73	0.70	4.51	223
	LIMIT	0.43	0.40	1.05	0.66	0.66	0.35	3.52	33

MODE		RET#	RET%	RBC-O	HFR	MFR	LFR	IRF
		10 ³ /uL	%	10 ⁶ /uL	%	%	%	%
Open	TARGET	0.0843	1.91	4.48	1.3	16.4	82.3	17.7
	LIMIT	0.0253	0.57	0.45	8.0	28.0	33.0	36.0
Closed	TARGET	0.0844	1.90	4.76	1.2	14.9	83.9	16.1
	LIMIT	0.0253	0.57	0.48	8.0	28.0	33.0	36.0

MODE		IG# ^{*2}	RET-He ^{*3}	
		10 ³ /uL	pg	fmol
Open	TARGET	0.93	23.9	1.48
	LIMIT	0.28	2.4	0.15
Closed	TARGET	0.93	23.8	1.48
	LIMIT	0.28	2.4	0.15

*1 : Available only on XT-4000i.

*2 : Available only on XT-4000i, XT-2000i, or XT-1800i installed the optional software, XT IG master.

*3 : Available only on XT-4000i, or XT-2000i installed the optional software, XT RET master.

Please refer to the help message provided in the assay CD for the handling of the e-CHECK.(XE).

Please make sure the lot number before reading the data from the assay CD.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915

< e-CHECK[®](XE) ASSAY SHEET >

For XT-Series

Lot No. 70220812

Exp. date 2017-04-16

L3 : Level 3

MODE		RBC 10 ⁶ /μL	HGB		HCT %	MCV fL	MCH		MCHC		PLT 10 ³ /μL
			g/dL	mmol/L			pg	fmol	g/dL	mmol/L	
Open	TARGET	5.28	16.5	10.2	47.3	89.5	31.3	1.93	34.9	21.6	546
	LIMIT	0.26	0.5	0.3	2.4	4.5	1.6	0.10	2.1	1.3	71
Closed	TARGET	5.24	16.6	10.3	47.2	90.1	31.7	1.97	35.2	21.8	510
	LIMIT	0.26	0.5	0.3	2.4	4.5	1.6	0.10	2.1	1.3	66

MODE		RDW-SD fL	RDW-CV %	PDW fL	PCT %	MPV fL	P-LCR %
Open	TARGET	44.2	14.0	8.9	0.54	9.8	17.4
	LIMIT	6.6	2.1	1.8	0.14	1.5	8.7
Closed	TARGET	43.6	14.0	8.8	0.50	9.8	17.2
	LIMIT	6.5	2.1	1.8	0.13	1.5	8.6

MODE		WBC 10 ³ /μL	WBC-D ^{*1} 10 ³ /μL	NEUT# 10 ³ /μL	LYMPH# 10 ³ /μL	MONO# 10 ³ /μL	EO# 10 ³ /μL	BASO# 10 ³ /μL	PLT-O 10 ³ /μL
Open	TARGET	18.81	17.54	9.92	4.74	2.09	2.06	13.43	568
	LIMIT	1.13	1.05	2.98	1.42	1.88	1.03	10.48	85
Closed	TARGET	18.12	18.01	9.52	4.64	1.98	1.98	12.90	566
	LIMIT	1.09	1.08	2.86	1.39	1.78	0.99	10.06	85

MODE		RET# 10 ⁶ /μL	RET% %	RBC-O 10 ⁶ /μL	HFR %	MFR %	LFR %	IRF %
Open	TARGET	0.0378	0.72	5.35	1.5	12.7	85.8	14.2
	LIMIT	0.0113	0.22	0.54	8.0	28.0	33.0	36.0
Closed	TARGET	0.0362	0.69	5.65	1.3	11.5	87.2	12.8
	LIMIT	0.0109	0.21	0.57	8.0	28.0	33.0	36.0

MODE		IG# ^{*2} 10 ³ /μL	RET-He ^{*3}	
			pg	fmol
Open	TARGET	2.73	24.9	1.55
	LIMIT	0.68	2.5	0.16
Closed	TARGET	2.61	24.7	1.53
	LIMIT	0.65	2.5	0.15

*1 : Available only on XT-4000i.

*2 : Available only on XT-4000i, XT-2000i, or XT-1800i installed the optional software, XT IG master.

*3 : Available only on XT-4000i, or XT-2000i installed the optional software, XT RET master.

Please refer to the help message provided in the assay CD for the handling of the e-CHECK (XE).

Please make sure the lot number before reading the data from the assay CD.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915