< *e* -CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 71560804 Exp. date 2017-08-27 **L1 : Level 1**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	2.35	5.9	3.7	17.9	76.2	25.1	1.56	33.0	20.5	59
LIMIT	0.12	0.2	0.1	1.1	3.8	1.3	0.08	2.0	1.2	24

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	43.4	15.8	9.1	0.06	9.7	17.0
LIMIT	6.5	2.4	3.6	0.03	1.5	14.5

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	3.00	2.92	1.11	1.15	0.32	0.26	0.16
LIMIT	0.30	0.29	0.33	0.46	0.29	0.13	0.12

Please refer to the help message provided in the assay CD-ROM for the handling of the *e* -CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915

< *e* -CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 71560805 Exp. date 2017-08-27 **L2 : Level 2**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	4.43	12.6	7.8	37.0	83.6	28.4	1.77	34.1	21.1	243
LIMIT	0.22	0.4	0.2	1.9	4.2	1.4	0.09	2.0	1.3	36

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	44.7	15.1	9.7	0.25	10.2	19.2
LIMIT	6.7	2.3	2.9	0.08	1.5	9.6

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	7.04	6.85	2.85	2.28	0.82	0.67	0.42
LIMIT	0.42	0.41	0.86	0.68	0.74	0.34	0.33

Please refer to the help message provided in the assay CD-ROM for the handling of the *e* -CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915