< *e*-CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 80710804 Exp. date 2018-06-03 **L1 : Level 1**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	2.40	5.7	3.5	17.6	73.4	23.8	1.47	32.4	20.1	58
LIMIT	0.12	0.2	0.1	1.1	3.7	1.2	0.07	1.9	1.2	23

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	45.9	17.4	9.8	0.06	9.9	18.5
LIMIT	6.9	2.6	3.9	0.03	1.5	15.7

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	2.91	2.82	1.11	1.07	0.32	0.25	0.16
LIMIT	0.29	0.28	0.33	0.43	0.29	0.13	0.12

Please refer to the help message provided in the assay CD-ROM for the handling of the *e*-CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915

< *e*-CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 80710805 Exp. date 2018-06-03 **L2 : Level 2**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	4.45	11.9	7.4	35.7	80.2	26.7	1.66	33.3	20.7	227
LIMIT	0.22	0.4	0.2	1.8	4.0	1.3	0.08	2.0	1.2	34

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	44.5	16.0	9.7	0.23	10.1	18.8
LIMIT	6.7	2.4	2.9	0.07	1.5	9.4

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	6.94	6.72	2.81	2.32	0.71	0.69	0.41
LIMIT	0.42	0.40	0.84	0.70	0.64	0.35	0.32

Please refer to the help message provided in the assay CD-ROM for the handling of the *e*-CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915