< *e* -CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 82950804 Exp. date 2019-01-13 **L1 : Level 1**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	2.32	5.6	3.5	17.5	75.5	24.1	1.50	32.0	19.9	57
LIMIT	0.12	0.2	0.1	1.1	3.8	1.2	0.08	1.9	1.2	23

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	53.5	19.8	8.7	0.06	9.8	17.3
LIMIT	8.0	3.0	3.5	0.03	1.5	14.7

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	2.98	2.89	1.06	1.21	0.30	0.26	0.15
LIMIT	0.30	0.29	0.32	0.48	0.27	0.13	0.12

Please refer to the help message provided in the assay CD-ROM for the handling of the *e* -CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8^{o}C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915

< *e* -CHECK(XS) ASSAY SHEET > **For XS-Series**

Lot No. 82950805 Exp. date 2019-01-13 **L2 : Level 2**

	RBC	HGB		HCT	MCV	MCH		MCHC		PLT
	10^6/uL	g/dL	mmol/L	%	fL	pg	fmol	g/dL	mmol/L	10^3/uL
TARGET	4.35	12.1	7.5	35.5	81.7	27.8	1.73	34.1	21.2	236
LIMIT	0.22	0.4	0.2	1.8	4.1	1.4	0.09	2.0	1.3	35

	RDW-SD	RDW-CV	PDW	PCT	MPV	P-LCR
	fL	%	fL	%	fL	%
TARGET	48.4	16.7	9.1	0.24	10.1	17.8
LIMIT	7.3	2.5	2.7	0.07	1.5	8.9

	WBC-D	WBC-C	NEUT#	LYMPH#	MONO#	EO#	BASO#
	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL	10^3/uL
TARGET	6.84	6.61	2.83	2.20	0.72	0.67	0.42
LIMIT	0.41	0.40	0.85	0.66	0.65	0.34	0.33

Please refer to the help message provided in the assay CD-ROM for the handling of the *e* -CHECK(XS).

Please make sure the lot number before reading the data from the assay CD-ROM.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8^{o}C) immediately after use.

U.S. Patents 5,432,089; 6,200,500; 6,221,668; 6,399,388; 6,403,377; 6,406,915